

---

## List of certificates relating to ATEX QA Notification and/ or IECEx QAR


---

Notification No.: DNV-2006-OSL-ATEX-0473Q / Presafe 15 ATEX 6595Q

IECEx QAR No.: NO/DNV/QAR06.0006/04 / NO/PRE/QAR15.0011/00

Manufacturer.: GP:50, 2770 Long Road, Grand Island, NY 14072 USA

Project No.: PRJC-58001-2008-PRC-USA

ATEX EC Type Examination Certificate	Description of ATEX equipment	Ex marking
DNV 07 ATEX 3697X Rev 2	Pressure and/or Temperature Transmitter Models 1**, 2**, 3**, with or without Temperature Output	 II 1 G Ex ia IIC T5 Ga -40°C to +85°C
IECEx certificate of Conformity	Description of IECEx equipment	Ex marking
IECEx DNV.07.0003X Issue 1	Pressure and/or Temperature Transmitter Models 1**, 2**, 3**, with or without Temperature Output	Ex ia IIC T5 Ga -40°C to +85°C
IECEx LCI 08.0056X Issue 0	Pressure Transducers Models 311*N*** and 370*N***	Ex nA IIC T5 Ex ic IIC T5 -40°C to +80°C

Date: 2016-11-08

Sign: Jacob Thumberger